

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW090819\
 Data File : VW012810.D
 Acq On : 09 Sep 2019 09:25
 Operator : SY/VA
 Sample : K4760-07
 Misc : 5.62G/5ML/MSVOA_W/SOIL
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 NORTH-SIDE-WALL-3

Quant Time: Sep 16 07:54:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090719S.M
 Quant Title : SW846 8260
 QLast Update : Sun Sep 08 05:04:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	264265	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	400613	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	315117	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	113477	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	143223	48.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.02%	
35) Dibromofluoromethane	7.88	113	115319	48.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.84%	
50) Toluene-d8	10.32	98	474153	50.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.62%	
62) 4-Bromofluorobenzene	12.62	95	149500	44.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.26%	
Target Compounds						
16) Acetone	4.14	43	79393	98.629	ug/l	96
17) Carbon Disulfide	4.37	76	12182	2.630	ug/l	100
52) Toluene	10.38	92	9099	1.476	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW090819\
Data File : VW012810.D
Acq On : 09 Sep 2019 09:25
Operator : SY/VA
Sample : K4760-07
Misc : 5.62G/5ML/MSVOA_W/SOIL
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
NORTH-SIDE-WALL-3

Quant Time: Sep 16 07:54:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W090719S.M
Quant Title : SW846 8260
QLast Update : Sun Sep 08 05:04:40 2019
Response via : Initial Calibration

