

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090920\
 Data File : VW016504.D
 Acq On : 09 Sep 2020 18:07
 Operator : SY/VA
 Sample : L3950-10
 Misc : 5.93G/10ML/MSVOA W/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFX01

Manual Integrations
 APPROVED

MMDadoda
 9/10/2020 5:17:56 PM

Quant Time: Sep 10 05:53:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM090320S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 10 05:09:14 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	675448	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	585940	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	259322	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	190713	26.61	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	106.44%
7) Chloroethane-d5	2.90	69	166139	25.82	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	103.28%
10) 1,1-Dichloroethene-d2	4.03	63	279782m	16.98	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.92%
20) 2-Butanone-d5	7.09	46	75152	29.36	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	58.72%
24) Chloroform-d	7.65	84	356273	22.52	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.08%
26) 1,2-Dichloroethane-d4	8.31	65	187426	21.66	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	86.64%
29) Benzene-d6	8.28	84	721537	24.22	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.88%
33) 1,2-Dichloropropane-d6	9.28	67	199266	22.77	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	91.08%
37) Toluene-d8	10.33	98	685823	24.22	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	96.88%
38) trans-1,3-Dichloropropene-	10.58	79	93458	21.17	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.68%
39) 2-Hexanone-d5	10.93	63	64712	36.18	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	72.36%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	150050	20.28	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	81.12%
61) 1,2-Dichlorobenzene-d4	13.85	152	215507	24.05	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.20%

Target Compounds					Ovalue
16) Methylene chloride	4.93	84	29625	2.810 ug/L	89
47) Tetrachloroethene	10.87	164	4328	0.608 ug/L	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090920\
Data File : VW016504.D
Acq On : 09 Sep 2020 18:07
Operator : SY/VA
Sample : L3950-10
Misc : 5.93G/10ML/MSVOA W/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFX01

Manual Integrations
APPROVED

MMDadoda
9/10/2020 5:17:56 PM

Quant Time: Sep 10 05:53:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM090320S.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 10 05:09:14 2020
Response via : Initial Calibration

