

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW090921\
 Data File : VW020004.D
 Acq On : 09 Sep 2021 15:39
 Operator : SY/VA
 Sample : VW0909SBL01
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0909SBL01

Quant Time: Sep 10 03:46:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W090721S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 08 12:18:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.958	168	175454	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.848	114	299559	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.634	117	281854	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.560	152	133010	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.311	65	99532	48.866	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.740%
35) Dibromofluoromethane	7.891	113	90321	45.913	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.820%
50) Toluene-d8	10.323	98	332861	45.518	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.040%
62) 4-Bromofluorobenzene	12.622	95	122516	45.280	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.560%
Target Compounds						
16) Acetone	4.135	43	3451	10.549	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW090921\
Data File : VW020004.D
Acq On : 09 Sep 2021 15:39
Operator : SY/VA
Sample : VW0909SBL01
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0909SBL01

Quant Time: Sep 10 03:46:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W090721S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 08 12:18:21 2021
Response via : Initial Calibration

