


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW091018\
Data File : VW005237.D
Acq On : 10 Sep 2018 12:23
Operator : SY/AP
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W091018S.M
Title : SW846 8260
Last Update : Tue Sep 11 03:48:23 2018



AutoFind: Scans 1758, 1759, 1760; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.2	34872	PASS
75	95	30	60	45.5	104669	PASS
95	95	100	100	100.0	230101	PASS
96	95	5	9	6.5	15007	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	88.4	203349	PASS
175	174	5	9	7.7	15680	PASS
176	174	95	101	95.9	195072	PASS
177	176	5	9	6.6	12901	PASS