

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW091118\
 Data File : VW005269.D
 Acq On : 11 Sep 2018 12:29
 Operator : SY/AP
 Sample : J4779-03
 Misc : 5.73G/5ML/MSVOA W/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SU-03-090618-B

Quant Time: Sep 11 13:06:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091018S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 03:48:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	401285	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	620796	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	546905	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	203827	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.32	65	145792	48.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.42%	
35) Dibromofluoromethane	7.89	113	138241	42.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.02%	
50) Toluene-d8	10.33	98	545991	33.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.46%	
62) 4-Bromofluorobenzene	12.63	95	153553	25.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	51.10%	
Target Compounds						
16) Acetone	4.14	43	15297	27.53	ug/l	99
20) Methylene Chloride	4.92	84	17862	1.11	ug/l	87
80) 1,3,5-Trimethylbenzene	12.95	105	15948	1.30	ug/l	98
84) 1,2,4-Trimethylbenzene	13.26	105	73794	5.99	ug/l	98
89) n-Butylbenzene	13.83	91	27833	2.17	ug/l #	75
95) Naphthalene	15.38	128	33723	4.70	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW091118\
Data File : VW005269.D
Acq On : 11 Sep 2018 12:29
Operator : SY/AP
Sample : J4779-03
Misc : 5.73G/5ML/MSVOA W/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SU-03-090618-B

Quant Time: Sep 11 13:06:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W091018S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 03:48:23 2018
Response via : Initial Calibration

