

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW091118\
 Data File : VW005271.D
 Acq On : 11 Sep 2018 13:20
 Operator : SY/AP
 Sample : J4809-05RE
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 MS-SLUDGE-V24129RE

Quant Time: Sep 11 14:24:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091018S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 03:48:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	386355	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	514684	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	389266	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	120343	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	111842	38.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.62%	
35) Dibromofluoromethane	7.89	113	116319	43.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.28%	
50) Toluene-d8	10.33	98	498591	36.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.20%	
62) 4-Bromofluorobenzene	12.63	95	115298	23.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	46.28%	
Target Compounds						
16) Acetone	4.13	43	44055	82.36	ug/l	93
20) Methylene Chloride	4.92	84	22684	2.58	ug/l	95
25) 2-Butanone	7.18	43	14755	17.49	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW091118\
Data File : VW005271.D
Acq On : 11 Sep 2018 13:20
Operator : SY/AP
Sample : J4809-05RE
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
MS-SLUDGE-V24129RE

Quant Time: Sep 11 14:24:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W091018S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 03:48:23 2018
Response via : Initial Calibration

