

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW091118\
 Data File : VW005284.D
 Acq On : 11 Sep 2018 18:52
 Operator : SY/AP
 Sample : J4776-03
 Misc : 6.61G/5ML/MSVOA W/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EO-02-090718-B

Quant Time: Sep 12 01:30:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091018S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 03:48:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	441612	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	697812	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	689856	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	355268	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.32	65	166430	50.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.06%	
35) Dibromofluoromethane	7.89	113	151986	41.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.16%	
50) Toluene-d8	10.33	98	711625	38.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.06%	
62) 4-Bromofluorobenzene	12.62	95	270642	40.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.12%	
Target Compounds						
16) Acetone	4.13	43	46136	75.46	ug/l	99
20) Methylene Chloride	4.93	84	28081	3.09	ug/l	97
30) Chloroform	7.68	83	28467	4.28	ug/l	97
80) 1,3,5-Trimethylbenzene	12.95	105	71992	3.37	ug/l	99
84) 1,2,4-Trimethylbenzene	13.26	105	275296	12.82	ug/l	99
86) p-Isopropyltoluene	13.51	119	28389	1.19	ug/l	98
95) Naphthalene	15.38	128	38688	3.09	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW091118\
Data File : VW005284.D
Acq On : 11 Sep 2018 18:52
Operator : SY/AP
Sample : J4776-03
Misc : 6.61G/5ML/MSVOA W/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EO-02-090718-B

Quant Time: Sep 12 01:30:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W091018S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 03:48:23 2018
Response via : Initial Calibration

