

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW091118\
 Data File : VW005286.D
 Acq On : 11 Sep 2018 19:43
 Operator : SY/AP
 Sample : J4838-05
 Misc : 5.67G/5ML/MSVOA W/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SUP-1B-G-090718-2

Quant Time: Sep 12 01:34:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091018S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 03:48:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	438666	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	701379	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	679603	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	315917	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	190915	58.35	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	116.70%	
35) Dibromofluoromethane	7.89	113	162087	44.12	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	88.24%	
50) Toluene-d8	10.33	98	703354	37.88	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	75.76%	
62) 4-Bromofluorobenzene	12.62	95	253836	37.38	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	74.76%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.12	43	108914	179.34	ug/l	100
20) Methylene Chloride	4.93	84	27669	3.04	ug/l	92
25) 2-Butanone	7.18	43	39470	41.20	ug/l	90
80) 1,3,5-Trimethylbenzene	12.95	105	30228	1.59	ug/l	95
84) 1,2,4-Trimethylbenzene	13.26	105	119431	6.25	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW091118\
Data File : VW005286.D
Acq On : 11 Sep 2018 19:43
Operator : SY/AP
Sample : J4838-05
Misc : 5.67G/5ML/MSVOA W/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SUP-1B-G-090718-2

Quant Time: Sep 12 01:34:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W091018S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 03:48:23 2018
Response via : Initial Calibration

