


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW091118\
 Data File : VW005264.D
 Acq On : 11 Sep 2018 09:47
 Operator : SY/AP
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091018S.M
 Title : SW846 8260
 Last Update : Tue Sep 11 03:48:23 2018



AutoFind: Scans 1758, 1759, 1760; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.1	36866	PASS
75	95	30	60	45.4	110589	PASS
95	95	100	100	100.0	243349	PASS
96	95	5	9	6.5	15863	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	89.1	216768	PASS
175	174	5	9	7.7	16606	PASS
176	174	95	101	96.3	208789	PASS
177	176	5	9	6.9	14322	PASS