

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW091218\
 Data File : VW005331.D
 Acq On : 12 Sep 2018 17:40
 Operator : SY/AP
 Sample : J4834-02
 Misc : 5.05G/5ML/MSVOA W/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C-1A

Quant Time: Sep 12 19:03:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091018S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 03:48:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	388481	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	593173	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	593656	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	301895	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	146471	50.55	ug/l	0.00
Spiked Amount				50.000		
Recovery				=	101.10%	
35) Dibromofluoromethane	7.89	113	137431	44.23	ug/l	0.00
Spiked Amount				50.000		
Recovery				=	88.46%	
50) Toluene-d8	10.33	98	615705	39.21	ug/l	0.00
Spiked Amount				50.000		
Recovery				=	78.42%	
62) 4-Bromofluorobenzene	12.63	95	221819	38.62	ug/l	0.00
Spiked Amount				50.000		
Recovery				=	77.24%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.13	43	3545	6.59	ug/l	90
20) Methylene Chloride	4.92	84	29107	4.27	ug/l	93
29) Tetrahydrofuran	7.54	42	925	1.62	ug/l #	89
59) 2-Hexanone	10.98	43	1637	1.05	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW091218\
Data File : VW005331.D
Acq On : 12 Sep 2018 17:40
Operator : SY/AP
Sample : J4834-02
Misc : 5.05G/5ML/MSVOA W/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C-1A

Quant Time: Sep 12 19:03:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W091018S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 03:48:23 2018
Response via : Initial Calibration

