

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW091218\
 Data File : VW005333.D
 Acq On : 12 Sep 2018 18:31
 Operator : SY/AP
 Sample : J4838-06
 Misc : 4.60G/5ML/MSVOA W/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SUP-1B-G-090718-3

Quant Time: Sep 13 01:11:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091018S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 03:48:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	353911	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	528171	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	484739	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	209449	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	139430	52.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.64%	
35) Dibromofluoromethane	7.89	113	124664	45.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.12%	
50) Toluene-d8	10.33	98	516722	36.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.92%	
62) 4-Bromofluorobenzene	12.62	95	164410	32.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	64.30%	
Target Compounds						
16) Acetone	4.14	43	5495	11.22	ug/l	Qvalue # 82
20) Methylene Chloride	4.92	84	34464	6.61	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW091218\
Data File : VW005333.D
Acq On : 12 Sep 2018 18:31
Operator : SY/AP
Sample : J4838-06
Misc : 4.60G/5ML/MSVOA W/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SUP-1B-G-090718-3

Quant Time: Sep 13 01:11:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W091018S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 03:48:23 2018
Response via : Initial Calibration

