

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW091218\
 Data File : VW005336.D
 Acq On : 12 Sep 2018 19:48
 Operator : SY/AP
 Sample : J4866-02
 Misc : 5.06G/5ML/MSVOA W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DPT-SB-E1-180911

Quant Time: Sep 13 01:16:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091018S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 11 03:48:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	333991	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	503395	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	487576	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	265744	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	134447	53.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.94%	
35) Dibromofluoromethane	7.89	113	124009	47.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.06%	
50) Toluene-d8	10.33	98	510935	38.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.68%	
62) 4-Bromofluorobenzene	12.62	95	188958	38.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.54%	
Target Compounds						
16) Acetone	4.15	43	12331	26.67	ug/l	Qvalue # 85
87) 1,3-Dichlorobenzene	13.51	146	13599	1.46	ug/l	92
88) 1,4-Dichlorobenzene	13.58	146	44118	4.84	ug/l	93
93) 1,2,4-Trichlorobenzene	15.15	180	9533	1.65	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW091218\
Data File : VW005336.D
Acq On : 12 Sep 2018 19:48
Operator : SY/AP
Sample : J4866-02
Misc : 5.06G/5ML/MSVOA W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DPT-SB-E1-180911

Quant Time: Sep 13 01:16:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W091018S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 11 03:48:23 2018
Response via : Initial Calibration

