


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW091218\
Data File : VW005317.D
Acq On : 12 Sep 2018 11:12
Operator : SY/AP
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091018S.M
Title : SW846 8260
Last Update : Tue Sep 11 03:48:23 2018



AutoFind: Scans 1758, 1759, 1760; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.5	39053	PASS
75	95	30	60	45.1	113680	PASS
95	95	100	100	100.0	251989	PASS
96	95	5	9	6.9	17434	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	87.4	220245	PASS
175	174	5	9	7.5	16499	PASS
176	174	95	101	96.7	212906	PASS
177	176	5	9	6.9	14647	PASS