

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW091319\
 Data File : VW012947.D
 Acq On : 13 Sep 2019 10:05
 Operator : SY/VA
 Sample : VW0913SBL01
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Sep 13 15:37:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM091219S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 13 15:34:49 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	428600	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	377724	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	180523	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	111834	25.95	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	103.80%
7) Chloroethane-d5	2.89	69	89487	25.24	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	100.96%
10) 1,1-Dichloroethene-d2	4.02	63	190601	18.49	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.96%
20) 2-Butanone-d5	7.07	46	77958	48.37	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	96.74%
24) Chloroform-d	7.65	84	254906	23.52	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.08%
26) 1,2-Dichloroethane-d4	8.31	65	141709	24.48	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.92%
29) Benzene-d6	8.27	84	480109	24.91	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	99.64%
33) 1,2-Dichloropropane-d6	9.27	67	152929	24.04	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.16%
37) Toluene-d8	10.32	98	454300	24.76	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	99.04%
38) trans-1,3-Dichloropropene-	10.57	79	65721	23.44	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	93.76%
39) 2-Hexanone-d5	10.92	63	55030	45.84	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.68%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	115444	24.04	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	96.16%
61) 1,2-Dichlorobenzene-d4	13.85	152	168398	24.49	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.96%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW091319\
Data File : VW012947.D
Acq On : 13 Sep 2019 10:05
Operator : SY/VA
Sample : VW0913SBL01
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Sep 13 15:37:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM091219S.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 13 15:34:49 2019
Response via : Initial Calibration

