

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW091519\
 Data File : VW013022.D
 Acq On : 15 Sep 2019 05:18
 Operator : SY/VA
 Sample : K4850-01
 Misc : 4.93G/5ML/MSVOA W/SOIL
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 AOC-7(5)

Quant Time: Sep 16 09:35:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091419S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 13 17:26:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	181326	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	253864	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	202052	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	76618	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	88581	52.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.78%	
35) Dibromofluoromethane	7.88	113	72572	52.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.36%	
50) Toluene-d8	10.32	98	303199	56.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.50%	
62) 4-Bromofluorobenzene	12.62	95	91470	47.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.38%	
Target Compounds						
					Qvalue	
16) Acetone	4.12	43	23244	48.185	ug/l	97
17) Carbon Disulfide	4.37	76	40353	16.178	ug/l	97
25) 2-Butanone	7.17	43	6740	12.657	ug/l	96
52) Toluene	10.39	92	5259	1.461	ug/l	100
68) m/p-Xylenes	11.84	106	4146	1.707	ug/l	86
73) Isopropylbenzene	12.46	105	133298	26.248	ug/l	98
78) n-propylbenzene	12.80	91	19474	3.260	ug/l	98
95) Naphthalene	15.37	128	8812	3.631	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW091519\
Data File : VW013022.D
Acq On : 15 Sep 2019 05:18
Operator : SY/VA
Sample : K4850-01
Misc : 4.93G/5ML/MSVOA W/SOIL
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
AOC-7-(5)

Quant Time: Sep 16 09:35:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W091419S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 13 17:26:01 2019
Response via : Initial Calibration

