

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW091718\
 Data File : VW005432.D
 Acq On : 17 Sep 2018 14:38
 Operator : SY/AP
 Sample : J4773-05
 Misc : 7.00G/5ML/MSVOA W/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 CL-01-091418-C

Quant Time: Sep 18 01:42:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091618S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 15 02:32:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	232514	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	375269	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	351630	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	145594	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	118021	48.90	ug/l	0.00
Spiked Amount			Recovery	=	97.80%	
35) Dibromofluoromethane	7.89	113	90104	42.82	ug/l	0.00
Spiked Amount			Recovery	=	85.64%	
50) Toluene-d8	10.33	98	384472	39.90	ug/l	0.00
Spiked Amount			Recovery	=	79.80%	
62) 4-Bromofluorobenzene	12.63	95	133124	37.68	ug/l	0.00
Spiked Amount			Recovery	=	75.36%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.14	43	64021	179.37	ug/l	98
17) Carbon Disulfide	4.39	76	7691	1.21	ug/l	100
18) Methyl Acetate	4.69	43	2614	2.25	ug/l	98
52) Toluene	10.40	92	11041	1.58	ug/l	99
68) m/p-Xylenes	11.84	106	11983	2.37	ug/l	94
69) o-Xylene	12.18	106	5800	1.23	ug/l	95
95) Naphthalene	15.38	128	175261	36.83	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW091718\
Data File : VW005432.D
Acq On : 17 Sep 2018 14:38
Operator : SY/AP
Sample : J4773-05
Misc : 7.00G/5ML/MSVOA W/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
CL-01-091418-C

Quant Time: Sep 18 01:42:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W091618S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 15 02:32:34 2018
Response via : Initial Calibration

