

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW091718\
 Data File : VW005444.D
 Acq On : 17 Sep 2018 19:43
 Operator : SY/AP
 Sample : J4960-01RE
 Misc : 5.17G/5ML/MSVOA W/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 S1RE

Quant Time: Sep 18 02:09:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091618S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 15 02:32:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	208413	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	321674	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	260996	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	85236	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	93665	43.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.60%	
35) Dibromofluoromethane	7.89	113	72076	39.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.92%	
50) Toluene-d8	10.33	98	288629	34.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.88%	
62) 4-Bromofluorobenzene	12.62	95	81912	27.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	54.10%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.15	43	73440	234.98	ug/l	98
18) Methyl Acetate	4.68	43	3128	3.01	ug/l	95
20) Methylene Chloride	4.93	84	19475	3.77	ug/l	96
25) 2-Butanone	7.19	43	25978	49.92	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW091718\
Data File : VW005444.D
Acq On : 17 Sep 2018 19:43
Operator : SY/AP
Sample : J4960-01RE
Misc : 5.17G/5ML/MSVOA W/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
S1RE

Quant Time: Sep 18 02:09:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W091618S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 15 02:32:34 2018
Response via : Initial Calibration

