

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW091718\
 Data File : VW005459.D
 Acq On : 18 Sep 2018 04:35
 Operator : SY/AP
 Sample : J4945-01
 Misc : 0.65G/5ML/MSVOA W/SOIL
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 RB47195RT

Quant Time: Sep 19 12:53:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091618S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 15 02:32:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	224924	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	347423	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	220724	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	69797	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	120691	51.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.38%	
35) Dibromofluoromethane	7.89	113	90300	46.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.72%	
50) Toluene-d8	10.33	98	295371	33.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	66.22%	
62) 4-Bromofluorobenzene	12.63	95	68478	20.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	41.88%	
Target Compounds						
16) Acetone	4.14	43	160585	496.03	ug/l	99
18) Methyl Acetate	4.68	43	24328	21.68	ug/l	100
20) Methylene Chloride	4.93	84	20951	3.73	ug/l	97
25) 2-Butanone	7.19	43	4199	7.48	ug/l	98
52) Toluene	10.40	92	48869	7.58	ug/l	99
70) Styrene	12.19	104	22919	4.70	ug/l #	60

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW091718\
Data File : VW005459.D
Acq On : 18 Sep 2018 04:35
Operator : SY/AP
Sample : J4945-01
Misc : 0.65G/5ML/MSVOA W/SOIL
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
RB47195RT

Quant Time: Sep 19 12:53:41 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091618S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 15 02:32:34 2018
Response via : Initial Calibration

