


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW091718\
Data File : VW005448.D
Acq On : 17 Sep 2018 22:40
Operator : SY/AP
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091618S.M
Title : SW846 8260
Last Update : Sat Sep 15 02:32:34 2018



AutoFind: Scans 1758, 1759, 1760; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	27858	PASS
75	95	30	60	50.6	72357	PASS
95	95	100	100	100.0	143018	PASS
96	95	5	9	6.7	9622	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	81.1	116000	PASS
175	174	5	9	7.8	9092	PASS
176	174	95	101	97.1	112594	PASS
177	176	5	9	6.6	7391	PASS