

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW091818\
 Data File : VW005475.D
 Acq On : 18 Sep 2018 13:33
 Operator : SY/AP
 Sample : VW0918SBL01
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0918SBL01

Quant Time: Sep 18 16:55:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091618S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 15 02:32:34 2018
 Response via : Initial Calibration

| Internal Standards | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-----------------------------|--------|------|----------|-------|--------|----------|
| 1) Pentafluorobenzene | 7.96 | 168 | 315504 | 50.00 | ug/l | 0.00 |
| 34) 1,4-Difluorobenzene | 8.85 | 114 | 486667 | 50.00 | ug/l | 0.00 |
| 63) Chlorobenzene-d5 | 11.63 | 117 | 465185 | 50.00 | ug/l | 0.00 |
| 72) 1,4-Dichlorobenzene-d4 | 13.57 | 152 | 229420 | 50.00 | ug/l | 0.00 |
| System Monitoring Compounds | | | | | | |
| 33) 1,2-Dichloroethane-d4 | 8.31 | 65 | 154079 | 47.05 | ug/l | 0.00 |
| Spiked Amount | 50.000 | | Recovery | = | 94.10% | |
| 35) Dibromofluoromethane | 7.89 | 113 | 130716 | 47.90 | ug/l | 0.00 |
| Spiked Amount | 50.000 | | Recovery | = | 95.80% | |
| 50) Toluene-d8 | 10.33 | 98 | 610014 | 48.82 | ug/l | 0.00 |
| Spiked Amount | 50.000 | | Recovery | = | 97.64% | |
| 62) 4-Bromofluorobenzene | 12.62 | 95 | 216628 | 47.28 | ug/l | 0.00 |
| Spiked Amount | 50.000 | | Recovery | = | 94.56% | |
| Target Compounds | | | | | | |
| 20) Methylene Chloride | 4.92 | 84 | 22363 | 1.29 | ug/l | 96 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW091818\
 Data File : VW005475.D
 Acq On : 18 Sep 2018 13:33
 Operator : SY/AP
 Sample : VW0918SBL01
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0918SBL01

Quant Time: Sep 18 16:55:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091618S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 15 02:32:34 2018
 Response via : Initial Calibration

