

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW091818\
 Data File : VW005498.D
 Acq On : 18 Sep 2018 23:26
 Operator : SY/AP
 Sample : J4959-01RE
 Misc : 5.05G/5ML/MSVOA W/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BACKFILL-SW2RE

Quant Time: Sep 19 06:56:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091618S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 15 02:32:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	258401	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	414675	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	398217	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	196269	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.32	65	124156	46.29	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	92.58%	
35) Dibromofluoromethane	7.90	113	96859	41.66	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	83.32%	
50) Toluene-d8	10.33	98	376506	35.36	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	70.72%	
62) 4-Bromofluorobenzene	12.63	95	161206	41.29	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	82.58%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	4.92	84	18839	1.51	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW091818\
Data File : VW005498.D
Acq On : 18 Sep 2018 23:26
Operator : SY/AP
Sample : J4959-01RE
Misc : 5.05G/5ML/MSVOA W/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BACKFILL-SW2RE

Quant Time: Sep 19 06:56:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W091618S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 15 02:32:34 2018
Response via : Initial Calibration

