

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW091818\
 Data File : VW005494.D
 Acq On : 18 Sep 2018 21:45
 Operator : SY/AP
 Sample : J4936-35RE
 Misc : 2.32G/5ML/MSVOA W/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 132-1-VOLTRE

Quant Time: Sep 19 06:01:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091618S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 15 02:32:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.96	168	394	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	488	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	19	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	0.00	152	0	0.00	ug/l	-13.57

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	229	55.99	ug/l	0.00
Spiked Amount						
			Recovery	=	111.98%	
35) Dibromofluoromethane	7.91	113	221	80.77	ug/l	0.02
Spiked Amount						
			Recovery	=	161.54%	
50) Toluene-d8	10.33	98	260	20.75	ug/l	0.00
Spiked Amount						
			Recovery	=	41.50%	
62) 4-Bromofluorobenzene	12.63	95	210	45.71	ug/l	0.01
Spiked Amount						
			Recovery	=	91.42%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	2.23	50	46	9.89	ug/l #	42
6) Chloroethane	2.94	64	102	39.50	ug/l #	43
11) Tert butyl alcohol	5.41	59	19	62.53	ug/l #	1
12) 1,1-Dichloroethene	4.06	96	68	22.31	ug/l #	1
13) Acrolein	4.07	56	25	140.26	ug/l #	12
14) Allyl chloride	4.68	41	72	12.69	ug/l #	67
16) Acetone	4.13	43	20	17.23	ug/l #	46
18) Methyl Acetate	4.66	43	27	13.74	ug/l #	54
19) Methyl tert-butyl Ether	5.43	73	48	5.80	ug/l #	52
20) Methylene Chloride	4.92	84	60	10.22	ug/l #	53
21) trans-1,2-Dichloroethene	5.62	96	22	6.04	ug/l #	1
23) Vinyl Acetate	6.30	43	25	3.93	ug/l #	74
25) 2-Butanone	7.15	43	21	21.35	ug/l #	49
27) cis-1,2-Dichloroethene	7.21	96	20	5.05	ug/l #	1
28) Bromochloromethane	7.26	49	40	15.20	ug/l #	12
29) Tetrahydrofuran	7.55	42	49	78.18	ug/l #	35
37) Ethyl Acetate	7.27	43	54	28.75	ug/l #	68
40) Benzene	8.16	78	22	1.64	ug/l #	51
41) Methacrylonitrile	7.49	41	19	16.77	ug/l #	30
43) Isopropyl Acetate	8.37	43	26	7.09	ug/l #	52
46) Dibromomethane	9.57	93	23	14.63	ug/l #	3
48) Methyl methacrylate	9.44	41	79	43.19	ug/l #	42
51) 4-Methyl-2-Pentanone	10.26	43	109	57.77	ug/l #	38
55) 1,1,2-Trichloroethane	10.71	97	44	18.83	ug/l #	15
56) Ethyl methacrylate	10.69	69	27	9.13	ug/l #	80
59) 2-Hexanone	10.99	43	81	63.02	ug/l #	27
67) Ethyl Benzene	11.82	91	22	31.02	ug/l #	44


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW091818\
Data File : VW005494.D
Acq On : 18 Sep 2018 21:45
Operator : SY/AP
Sample : J4936-35RE
Misc : 2.32G/5ML/MSVOA W/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
132-1-VOLTRE

Quant Time: Sep 19 06:01:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W091618S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 15 02:32:34 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 7.96 min Scan# 993

Delta R.T. -0.00 min

Lab File: VW005494.D

Acq: 18 Sep 2018 21:45

Instrument :

MSVOA_W

ClientSampleId :


132-1-VOLTRE

Tot Ion:168 Resp: 394

Ion Ratio Lower Upper


168 100

99 0.0 45.0 67.6#





Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VW



Time--> 7.90 7.95 8.00



#3

Chloromethane

Concen: 9.89 ug/l

RT: 2.23 min Scan# 53

Delta R.T. 0.01 min

Lab File: VW005494.D


Acq: 18 Sep 2018 21:45

Tgt Ion: 50 Resp: 46

Ion Ratio Lower Upper


50 100

52 0.0 25.8 38.6#





Abundance Ion 49.90 (49.60 to 50.60): VW

Ion 51.90 (51.60 to 52.60): VW



Time--> 2.20 2.22 2.24





#6

Chloroethane

Concen: 39.50 ug/l

RT: 2.94 min Scan# 170

Delta R.T. 0.02 min

Lab File: VW005494.D

Acq: 18 Sep 2018 21:45

Instrument :

MSVOA_W

ClientSampleId :





132-1-VOLTRE

Tot Ion: 64 Resp: 102

Ion Ratio Lower Upper

64 100

66 0.0 25.2 37.8#



#11

Tert butyl alcohol

Concen: 62.53 ug/l

RT: 5.41 min Scan# 575

Delta R.T. 0.20 min

Lab File: VW005494.D




Acq: 18 Sep 2018 21:45


Tgt Ion: 59 Resp: 19

Ion Ratio Lower Upper

59 100

57 157.9 8.6 13.0#









#12
 1,1-Dichloroethene
 Concen: 22.31 ug/l
 RT: 4.06 min Scan# 354
 Delta R.T. 0.01 min
 Lab File: VW005494.D
 Acq: 18 Sep 2018 21:45




Instrument :
 MSVOA_W
 ClientSampleId :
 132-1-VOLTRE


Tot Ion	Ratio	Lower	Upper
96	100		
61	0.0	135.7	203.5#
98	0.0	48.6	73.0#



#13
 Acrolein
 Concen: 140.26 ug/l
 RT: 4.07 min Scan# 355
 Delta R.T. 0.16 min
 Lab File: VW005494.D
 Acq: 18 Sep 2018 21:45

Tgt Ion	Ratio	Lower	Upper
56	100		
55	0.0	59.8	89.8#





#14

Allvl chloride

Concen: 12.69 ug/l

RT: 4.68 min Scan# 455

Delta R.T. -0.01 min

Lab File: VW005494.D

Acq: 18 Sep 2018 21:45

Instrument :

MSVOA_W

ClientSampled :

132-1-VOLTRE


Tot Ion: 41 Resp: 72

Ion Ratio Lower Upper

41 100

39 90.3 58.6 88.0#




76 0.0 27.4 41.0#



Abundance Ion 41.00 (40.70 to 41.70): VW

Ion 39.00 (38.70 to 39.70): VW

Ion 75.90 (75.60 to 76.60): VW



#16

Acetone

Concen: 17.23 ug/l

RT: 4.13 min Scan# 366

Delta R.T. -0.01 min

Lab File: VW005494.D


Acq: 18 Sep 2018 21:45

Tgt Ion: 43 Resp: 20

Ion Ratio Lower Upper



43 100


58 0.0 22.8 34.2#



Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 58.00 (57.70 to 58.70): VW









#18
Methyl Acetate
Concen: 13.74 ug/l
RT: 4.66 min Scan# 452
Delta R.T. -0.02 min
Lab File: VW005494.D
Acq: 18 Sep 2018 21:45




Instrument :
MSVOA_W
ClientSampled :
132-1-VOLTRE


Tot Ion: 43 Resp: 27
Ion Ratio Lower Upper
43 100
74 0.0 17.7 26.5#



#19
Methyl tert-butyl Ether
Concen: 5.80 ug/l
RT: 5.43 min Scan# 578
Delta R.T. -0.01 min
Lab File: VW005494.D
Acq: 18 Sep 2018 21:45

Tgt Ion: 73 Resp: 48
Ion Ratio Lower Upper
73 100
57 0.0 18.6 28.0#





#20

Methylene Chloride

Concen: 10.22 ug/l

RT: 4.92 min Scan# 494

Delta R.T. -0.01 min

Lab File: VW005494.D

Acq: 18 Sep 2018 21:45

Instrument :
MSVOA_W
ClientSampled :
132-1-VOLTRE

Tot Ion: 84 Resp: 60


Ion Ratio Lower Upper

84 100

49 153.6 99.3 148.9#

51 0.0 31.1 46.7#

86 0.0 49.1 73.7#





Abundance Ion 83.90 (83.60 to 84.60): VW

Ion 48.90 (48.60 to 49.60): VW

Ion 50.90 (50.60 to 51.60): VW

Ion 85.90 (85.60 to 86.60): VW



#21

trans-1,2-Dichloroethene

Concen: 6.04 ug/l

RT: 5.62 min Scan# 609

Delta R.T. 0.18 min

Lab File: VW005494.D

Acq: 18 Sep 2018 21:45



Tgt Ion: 96 Resp: 22

Ion Ratio Lower Upper

96 100

61 0.0 116.7 175.1#



98 0.0 51.0 76.6#




Abundance Ion 95.90 (95.60 to 96.60): VW

Ion 60.90 (60.60 to 61.60): VW

Ion 97.90 (97.60 to 98.60): VW





#23

Vinyl Acetate

Concen: 3.93 ug/l

RT: 6.30 min Scan# 721

Delta R.T. 0.03 min

Lab File: VW005494.D

Acq: 18 Sep 2018 21:45

Instrument :

MSVOA_W

ClientSampleId :


132-1-VOLTRE

Tot Ion: 43 Resp: 25

Ion Ratio Lower Upper

43 100

86 0.0 7.5 11.3#





Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 86.00 (85.70 to 86.70): VW

6.30

6.28 6.29 6.30 6.31 6.32

Time-->



#25

2-Butanone

Concen: 21.35 ug/l

RT: 7.15 min Scan# 861

Delta R.T. -0.03 min

Lab File: VW005494.D


Acq: 18 Sep 2018 21:45

Tgt Ion: 43 Resp: 21

Ion Ratio Lower Upper

43 100

72 0.0 21.1 31.7#




Abundance Ion 43.00 (42.70 to 43.70): VW


Ion 72.00 (71.70 to 72.70): VW

7.15

7.13 7.14 7.15 7.16 7.17

Time-->






#27



cis-1,2-Dichloroethene
 Concen: 5.05 ug/l
 RT: 7.21 min Scan# 871
 Delta R.T. 0.03 min
 Lab File: VW005494.D
 Acq: 18 Sep 2018 21:45

Instrument :
 MSVOA_W
 ClientSampleId :
 132-1-VOLTRE

Tot Ion	Ratio	Lower	Upper
96	100		
61	0.0	0.0	316.0
98	0.0	0.0	130.8





Abundance Ion 95.90 (95.60 to 96.60): VW
 Ion 60.90 (60.60 to 61.60): VW
 Ion 97.90 (97.60 to 98.60): VW





#28


Bromochloromethane
 Concen: 15.20 ug/l
 RT: 7.26 min Scan# 878
 Delta R.T. -0.26 min
 Lab File: VW005494.D
 Acq: 18 Sep 2018 21:45

Tgt Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	1.0
130	0.0	61.4	92.2#



Abundance Ion 48.90 (48.60 to 49.60): VW
 Ion 128.80 (128.50 to 129.50): V
 Ion 129.80 (129.50 to 130.50): V









#29
Tetrahydrofuran
Concen: 78.18 ug/l
RT: 7.55 min Scan# 926
Delta R.T. 0.01 min
Lab File: VW005494.D
Acq: 18 Sep 2018 21:45

Instrument :
MSVOA_W
ClientSampleId :
132-1-VOLTRE

Tot Ion: 42 Resp: 49
Ion Ratio Lower Upper
42 100
72 0.0 33.2 49.8#
71 0.0 31.0 46.6#




Abundance Ion 42.00 (41.70 to 42.70): VW
Ion 72.00 (71.70 to 72.70): VW
Ion 71.00 (70.70 to 71.70): VW






#33
1,2-Dichloroethane-d4
Concen: 55.99 ug/l
RT: 8.31 min Scan# 1051
Delta R.T. -0.00 min
Lab File: VW005494.D
Acq: 18 Sep 2018 21:45

Tgt Ion: 65 Resp: 229
Ion Ratio Lower Upper
65 100
67 10.5 0.0 100.6



Abundance Ion 64.90 (64.60 to 65.60): VW
Ion 66.90 (66.60 to 67.60): VW







#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 8.85 min Scan# 1139
 Delta R.T. -0.00 min
 Lab File: VW005494.D
 Acq: 18 Sep 2018 21:45

Instrument :
 MSVOA_W
 ClientSampleId :
 132-1-VOLTRE



Tot Ion:114 Resp: 488
 Ion Ratio Lower Upper
 114 100
 63 26.5 0.0 43.4
 88 16.6 0.0 31.8



Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VW
 Ion 87.90 (87.60 to 88.60): VW




Time--> 8.80 8.85 8.90




#35
 Dibromofluoromethane
 Concen: 80.77 ug/l
 RT: 7.91 min Scan# 985
 Delta R.T. 0.02 min
 Lab File: VW005494.D
 Acq: 18 Sep 2018 21:45


Tgt Ion:113 Resp: 221
 Ion Ratio Lower Upper
 113 100
 111 79.6 82.3 123.5#
 192 0.0 16.5 24.7#




Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V



Time--> 7.85 7.90





#37

Ethyl Acetate

Concen: 28.75 ug/l

RT: 7.27 min Scan# 880

Delta R.T. 0.01 min

Lab File: VW005494.D

Acq: 18 Sep 2018 21:45

Instrument :

MSVOA_W

ClientSampleId :

132-1-VOLTRE


Tot Ion: 43 Resp: 54

Ion Ratio Lower Upper

43 100

61 0.0 11.3 16.9#




70 0.0 8.2 12.4#



Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 60.90 (60.60 to 61.60): VW

Ion 70.00 (69.70 to 70.70): VW



#40

Benzene

Concen: 1.64 ug/l

RT: 8.16 min Scan# 1027

Delta R.T. -0.16 min

Lab File: VW005494.D


Acq: 18 Sep 2018 21:45

Tgt Ion: 78 Resp: 22

Ion Ratio Lower Upper



78 100


77 0.0 19.3 28.9#



Abundance Ion 78.00 (77.70 to 78.70): VW

Ion 77.00 (76.70 to 77.70): VW









#41
Methacrylonitrile
Concen: 16.77 ug/l
RT: 7.49 min Scan# 917
Delta R.T. -0.00 min
Lab File: VW005494.D
Acq: 18 Sep 2018 21:45




Instrument :
MSVOA_W
ClientSampleId :
132-1-VOLTRE


Tot Ion:	41	Resp:	19
Ion	Ratio	Lower	Upper
41	100		
39	100.0	44.6	67.0#
67	0.0	53.6	80.4#
52	0.0	24.9	37.3#



#43
Isopropyl Acetate
Concen: 7.09 ug/l
RT: 8.37 min Scan# 1060
Delta R.T. -0.07 min
Lab File: VW005494.D
Acq: 18 Sep 2018 21:45

Tgt Ion:	43	Resp:	26
Ion	Ratio	Lower	Upper
43	100		
61	0.0	23.8	35.6#
87	0.0	10.1	15.1#





#46

Dibromomethane

Concen: 14.63 ug/l

RT: 9.57 min Scan# 1257

Delta R.T. 0.10 min

Lab File: VW005494.D

Acq: 18 Sep 2018 21:45

Instrument :

MSVOA_W

ClientSampleId :

132-1-VOLTRE


Tot Ion: 93 Resp: 23

Ion Ratio Lower Upper

93 100

95 0.0 65.4 98.0#




174 0.0 83.4 125.2#



Abundance Ion 92.80 (92.50 to 93.50): VW

Ion 94.80 (94.50 to 95.50): VW

Ion 173.70 (173.40 to 174.40): V



#48

Methyl methacrylate

Concen: 43.19 ug/l

RT: 9.44 min Scan# 1236

Delta R.T. -0.01 min

Lab File: VW005494.D

Acq: 18 Sep 2018 21:45


Tgt Ion: 41 Resp: 79

Ion Ratio Lower Upper

41 100

69 0.0 67.5 101.3#



39 60.8 45.5 68.3




Abundance Ion 41.00 (40.70 to 41.70): VW

Ion 69.00 (68.70 to 69.70): VW

Ion 39.00 (38.70 to 39.70): VW





#50

Toluene-d8

Concen: 20.75 ug/l

RT: 10.33 min Scan# 1382

Delta R.T. -0.00 min

Lab File: VW005494.D

Acq: 18 Sep 2018 21:45

Instrument :

MSVOA_W

ClientSampleId :


132-1-VOLTRE

Tot Ion: 98 Resp: 260

Ion Ratio Lower Upper

98 100

100 46.9 51.4 77.0#



Abundance Ion 98.00 (97.70 to 98.70): VW

Ion 100.00 (99.70 to 100.70): VW

10.33



150

100

50

0

Time-->



#51

4-Methyl-2-Pentanone

Concen: 57.77 ug/l

RT: 10.26 min Scan# 1371

Delta R.T. 0.04 min

Lab File: VW005494.D


Acq: 18 Sep 2018 21:45

Tgt Ion: 43 Resp: 109

Ion Ratio Lower Upper

43 100

58 0.0 29.9 44.9#



Abundance Ion 43.00 (42.70 to 43.70): VW

Ion 58.00 (57.70 to 58.70): VW

10.26

80


60


40

20

0

Time-->









#55
 1,1,2-Trichloroethane
 Concen: 18.83 ug/l
 RT: 10.71 min Scan# 1445
 Delta R.T. -0.08 min
 Lab File: VW005494.D
 Acq: 18 Sep 2018 21:45




Instrument :
 MSVOA_W
 ClientSampleId :
 132-1-VOLTRE


Tot Ion: 97 Resp: 44
 Ion Ratio Lower Upper
 97 100
 83 0.0 67.8 101.8#
 85 0.0 45.4 68.0#
 99 0.0 51.5 77.3#



#56
 Ethyl methacrylate
 Concen: 9.13 ug/l
 RT: 10.69 min Scan# 1442
 Delta R.T. 0.04 min
 Lab File: VW005494.D
 Acq: 18 Sep 2018 21:45

Tgt Ion: 69 Resp: 27
 Ion Ratio Lower Upper
 69 100
 41 74.1 59.3 88.9
 39 74.1 30.2 45.4#









#59
2-Hexanone
Concen: 63.02 ug/l
RT: 10.99 min Scan# 1490
Delta R.T. 0.01 min
Lab File: VW005494.D
Acq: 18 Sep 2018 21:45




Instrument :
MSVOA_W
ClientSampled :
132-1-VOLTRE


Tot Ion: 43 Resp: 81
Ion Ratio Lower Upper
43 100
58 0.0 25.5 76.5#



#62
4-Bromofluorobenzene
Concen: 45.71 ug/l
RT: 12.63 min Scan# 1760
Delta R.T. 0.01 min
Lab File: VW005494.D
Acq: 18 Sep 2018 21:45

Tgt Ion: 95 Resp: 210
Ion Ratio Lower Upper
95 100
174 0.0 0.0 161.2
176 0.0 0.0 156.6









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.64 min Scan# 1597
Delta R.T. -0.00 min
Lab File: VW005494.D
Acq: 18 Sep 2018 21:45

Instrument :
MSVOA_W
ClientSampleId :
132-1-VOLTRE

Tot Ion:117 Resp: 19
Ion Ratio Lower Upper
117 100
82 0.0 45.9 68.9#
119 0.0 24.8 37.2#




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VW
Ion 118.90 (118.60 to 119.60): V






#67
Ethyl Benzene
Concen: 31.02 ug/l
RT: 11.82 min Scan# 1626
Delta R.T. 0.08 min
Lab File: VW005494.D
Acq: 18 Sep 2018 21:45

Tgt Ion: 91 Resp: 22
Ion Ratio Lower Upper
91 100
106 0.0 24.2 36.4#



Abundance Ion 91.00 (90.70 to 91.70): VW
Ion 106.00 (105.70 to 106.70): V





#72
 1,4-Dichlorobenzene-d4
 Concen: 0.00 ug/l
 Expected RT: 13.57 min
 Lab File: VW005494.D
 Acq: 18 Sep 2018 21:45

Instrument :
 MSVOA_W
 ClientSampleId :
 132-1-VOLTRE

Tot Ion: 152
 Sig Exp Ratio
 152 100
 115 59.0
 150 172.8

