

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW091819\
 Data File : VW013131.D
 Acq On : 19 Sep 2019 00:14
 Operator : SY/VA
 Sample : VW0918SBL02
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0918SBL02

Quant Time: Sep 19 09:30:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091419S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 13 17:26:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	402438	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	596712	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	498264	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	211965	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	169207	45.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.04%	
35) Dibromofluoromethane	7.88	113	163833	50.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.20%	
50) Toluene-d8	10.32	98	697651	55.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.12%	
62) 4-Bromofluorobenzene	12.62	95	215567	47.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.64%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW091819\
Data File : VW013131.D
Acq On : 19 Sep 2019 00:14
Operator : SY/VA
Sample : VW0918SBL02
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0918SBL02

Quant Time: Sep 19 09:30:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W091419S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 13 17:26:01 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.95 min Scan# 991

Delta R.T. -0.00 min

Lab File: VW013131.D

Acq: 19 Sep 2019 00:14

Instrument :

MSVOA_W

ClientSampleId :


VW0918SBL02

Tot Ion:168 Resp: 402438

Ion Ratio Lower Upper

168 100

99 50.4 46.2 69.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VW

7.95

200000

150000



100000

50000

0

7.90 8.00

Time-->



#33

1,2-Dichloroethane-d4

Concen: 45.525 ug/l

RT: 8.30 min Scan# 1049

Delta R.T. -0.01 min

Lab File: VW013131.D


Acq: 19 Sep 2019 00:14

Tgt Ion: 65 Resp: 169207

Ion Ratio Lower Upper

65 100

67 53.8 0.0 101.2



Abundance Ion 64.90 (64.60 to 65.60): VW

Ion 66.90 (66.60 to 67.60): VW

8.30

80000

60000


40000


20000

0

8.20 8.25 8.30 8.35 8.40

Time-->









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.84 min Scan# 1138
 Delta R.T. -0.00 min
 Lab File: VW013131.D
 Acq: 19 Sep 2019 00:14

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0918SBL02

Tot Ion	Ratio	Lower	Upper
114	100		
63	19.1	0.0	43.4
88	15.4	0.0	31.6




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): V
 Ion 87.90 (87.60 to 88.60): V






#35
 Dibromofluoromethane
 Concen: 50.600 ug/l
 RT: 7.88 min Scan# 980
 Delta R.T. -0.00 min
 Lab File: VW013131.D
 Acq: 19 Sep 2019 00:14

Tgt Ion	Ratio	Lower	Upper
113	100		
111	101.7	82.8	124.2
192	24.0	17.8	26.8



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 55.063 ug/l

RT: 10.32 min Scan# 1381

Delta R.T. -0.00 min

Lab File: VW013131.D

Acq: 19 Sep 2019 00:14

Instrument :

MSVOA_W

ClientSampleId :


VW0918SBL02

Tot Ion: 98 Resp: 697651

Ion Ratio Lower Upper




98 100

100 65.7 52.1 78.1



Abundance Ion 98.00 (97.70 to 98.70): VW

Ion 100.00 (99.70 to 100.70): VW



#62

4-Bromofluorobenzene

Concen: 47.316 ug/l

RT: 12.62 min Scan# 1757

Delta R.T. -0.00 min

Lab File: VW013131.D

Acq: 19 Sep 2019 00:14


Tgt Ion: 95 Resp: 215567

Ion Ratio Lower Upper

95 100

174 87.3 0.0 170.4



176 85.0 0.0 164.6




Abundance Ion 94.90 (94.60 to 95.60): VW

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.63 min Scan# 1595
Delta R.T. -0.00 min
Lab File: VW013131.D
Acq: 19 Sep 2019 00:14

Instrument :
MSVOA_W
ClientSampleId :
VW0918SBL02

Tot Ion	Ratio	Lower	Upper
117	100		
82	56.0	46.3	69.5
119	32.2	26.2	39.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.56 min Scan# 1912
Delta R.T. 0.01 min
Lab File: VW013131.D
Acq: 19 Sep 2019 00:14

Tgt Ion	Ratio	Lower	Upper
152	100		
115	56.0	28.8	86.4
150	156.4	0.0	340.2

