

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW091819\
 Data File : VW013147.D
 Acq On : 19 Sep 2019 07:33
 Operator : SY/VA
 Sample : K4825-08
 Misc : 4.75G/5ML/MSVOA W/SOIL
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 CSA-4-2-0.5

Quant Time: Sep 19 09:53:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W091419S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 13 17:26:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	348655	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	532824	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	345464	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.55	152	84610	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	156506	48.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.20%	
35) Dibromofluoromethane	7.88	113	151847	52.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.04%	
50) Toluene-d8	10.32	98	559531	49.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.92%	
62) 4-Bromofluorobenzene	12.62	95	120127	29.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	59.06%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW091819\
Data File : VW013147.D
Acq On : 19 Sep 2019 07:33
Operator : SY/VA
Sample : K4825-08
Misc : 4.75G/5ML/MSVOA W/SOIL
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
CSA-4-2-0.5

Quant Time: Sep 19 09:53:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W091419S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 13 17:26:01 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.95 min Scan# 991

Delta R.T. -0.00 min

Lab File: VW013147.D

Acq: 19 Sep 2019 07:33

Instrument :

MSVOA_W

ClientSampleId :


CSA-4-2-0.5

Tot Ion:168 Resp: 348655

Ion Ratio Lower Upper




168 100

99 51.3 46.2 69.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VW



#33

1,2-Dichloroethane-d4

Concen: 48.603 ug/l

RT: 8.31 min Scan# 1050

Delta R.T. -0.00 min

Lab File: VW013147.D


Acq: 19 Sep 2019 07:33

Tgt Ion: 65 Resp: 156506

Ion Ratio Lower Upper



65 100


67 53.0 0.0 101.2



Abundance Ion 64.90 (64.60 to 65.60): VW

Ion 66.90 (66.60 to 67.60): VW





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.84 min Scan# 1138

Delta R.T. -0.00 min

Lab File: VW013147.D

Acq: 19 Sep 2019 07:33

Instrument :

MSVOA_W

ClientSampleId :

CSA-4-2-0.5


Tot Ion:114 Resp: 532824

Ion Ratio Lower Upper

114 100

63 19.5 0.0 43.4




88 14.8 0.0 31.6



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VW

Ion 87.90 (87.60 to 88.60): VW



#35

Dibromofluoromethane

Concen: 52.521 ug/l

RT: 7.88 min Scan# 980

Delta R.T. -0.00 min

Lab File: VW013147.D

Acq: 19 Sep 2019 07:33


Tgt Ion:113 Resp: 151847

Ion Ratio Lower Upper

113 100

111 103.8 82.8 124.2



192 24.9 17.8 26.8




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.457 ug/l

RT: 10.32 min Scan# 1381

Delta R.T. -0.00 min

Lab File: VW013147.D

Acq: 19 Sep 2019 07:33

Instrument :

MSVOA_W

ClientSampleId :





CSA-4-2-0.5

Tot Ion: 98 Resp: 559531

Ion Ratio Lower Upper

98 100

100 66.1 52.1 78.1



#62

4-Bromofluorobenzene

Concen: 29.529 ug/l

RT: 12.62 min Scan# 1757

Delta R.T. -0.00 min

Lab File: VW013147.D

Acq: 19 Sep 2019 07:33




Tgt Ion: 95 Resp: 120127


Ion Ratio Lower Upper

95 100

174 87.3 0.0 170.4

176 86.6 0.0 164.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.63 min Scan# 1595
Delta R.T. -0.00 min
Lab File: VW013147.D
Acq: 19 Sep 2019 07:33

Instrument :
MSVOA_W
ClientSampleId :
CSA-4-2-0.5

Tot Ion:117 Resp: 345464
Ion Ratio Lower Upper
117 100
82 54.7 46.3 69.5
119 31.6 26.2 39.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.55 min Scan# 1911
Delta R.T. -0.00 min
Lab File: VW013147.D
Acq: 19 Sep 2019 07:33

Tgt Ion:152 Resp: 84610
Ion Ratio Lower Upper
152 100
115 55.6 28.8 86.4
150 157.4 0.0 340.2

