

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092221\
 Data File : VW020121.D
 Acq On : 22 Sep 2021 21:18
 Operator : SY/VA
 Sample : M3837-06RE
 Misc : 5.86g/10mL/MSVOA_W/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW7T6RE

Quant Time: Sep 23 01:27:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092121SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Sep 23 01:24:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	199143	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	180151	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	53323	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	99573	24.441	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.760%
7) Chloroethane-d5	2.892	69	79563	25.924	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	103.680%
11) 1,1-Dichloroethene-d2	4.019	63	106610	16.713	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	66.840%
21) 2-Butanone-d5	7.080	46	30884	49.125	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.260%
24) Chloroform-d	7.653	84	150340	24.005	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	96.000%
26) 1,2-Dichloroethane-d4	8.305	65	85534	25.375	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	101.520%
32) Benzene-d6	8.281	84	275564	23.976	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.920%
36) 1,2-Dichloropropane-d6	9.275	67	83989	25.250	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.000%
41) Toluene-d8	10.329	98	231017	22.091	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.360%
43) trans-1,3-Dichloroprop...	10.579	79	31633	22.250	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.000%
47) 2-Hexanone-d5	10.927	63	21119	45.880	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.760%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	61067	23.801	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	95.200%
66) 1,2-Dichlorobenzene-d4	13.853	152	51209	25.612	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	102.440%
Target Compounds						
					Qvalue	
13) Acetone	4.129	43	1878	3.488	ug/L	52
16) Methylene chloride	4.916	84	32300	9.192	ug/L	99
42) Toluene	10.390	91	3753	0.325	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092221\
Data File : VW020121.D
Acq On : 22 Sep 2021 21:18
Operator : SY/VA
Sample : M3837-06RE
Misc : 5.86g/10mL/MSVOA_W/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW7T6RE

Quant Time: Sep 23 01:27:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092121SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Sep 23 01:24:43 2021
Response via : Initial Calibration

