

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092221\
 Data File : VW020123.D
 Acq On : 22 Sep 2021 22:06
 Operator : SY/VA
 Sample : M3837-08
 Misc : 4.77g/10mL/MSVOA_W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW7W2

Quant Time: Sep 23 01:27:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092121SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Sep 23 01:24:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.847	114	266280	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	255399	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	95138	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	120262	22.076	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	88.320%	
7) Chloroethane-d5	2.891	69	96654	23.552	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	94.200%	
11) 1,1-Dichloroethene-d2	4.025	63	131544	15.423	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	61.680%	
21) 2-Butanone-d5	7.079	46	35372	42.078	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	84.160%	
24) Chloroform-d	7.652	84	186898	22.318	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	89.280%	
26) 1,2-Dichloroethane-d4	8.305	65	104088	23.094	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	92.360%	
32) Benzene-d6	8.280	84	348853	21.410	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	85.640%	
36) 1,2-Dichloropropane-d6	9.274	67	104137	22.084	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	88.320%	
41) Toluene-d8	10.329	98	304648	20.549	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	82.200%	
43) trans-1,3-Dichloroprop...	10.579	79	40512	20.100	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	80.400%	
47) 2-Hexanone-d5	10.926	63	25205	38.624	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	77.240%	
56) 1,1,2,2-Tetrachloroeth...	12.694	84	75733	20.820	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	83.280%	
66) 1,2-Dichlorobenzene-d4	13.853	152	85818	24.056	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	96.240%	
Target Compounds					Qvalue	
13) Acetone	4.129	43	7457	10.358	ug/L	99
16) Methylene chloride	4.915	84	25567	5.441	ug/L	96
42) Toluene	10.390	91	5397	0.330	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092221\
Data File : VW020123.D
Acq On : 22 Sep 2021 22:06
Operator : SY/VA
Sample : M3837-08
Misc : 4.77g/10mL/MSVOA_W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW7W2

Quant Time: Sep 23 01:27:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092121SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Sep 23 01:24:43 2021
Response via : Initial Calibration

