

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW092319\
 Data File : VW013232.D
 Acq On : 23 Sep 2019 15:38
 Operator : SY/VA
 Sample : K4918-08RE
 Misc : 3.47G/5ML/MSVOA W/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-9RE

Quant Time: Sep 24 07:28:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W092019S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 20 15:58:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	290577	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	436642	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	268306	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	60777	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	127898	53.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.86%	
35) Dibromofluoromethane	7.88	113	126137	52.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.04%	
50) Toluene-d8	10.32	98	441308	43.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.82%	
62) 4-Bromofluorobenzene	12.62	95	90443	26.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	52.68%	
Target Compounds						
16) Acetone	4.12	43	19661	36.763	ug/l	99
20) Methylene Chloride	4.91	84	58165	19.493	ug/l	93
64) Tetrachloroethene	10.86	164	3475	1.707	ug/l	90
68) m/p-Xylenes	11.83	106	9098	2.352	ug/l	98
69) o-Xylene	12.16	106	3869	1.076	ug/l	80
95) Naphthalene	15.36	128	10270	4.947	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW092319\
Data File : VW013232.D
Acq On : 23 Sep 2019 15:38
Operator : SY/VA
Sample : K4918-08RE
Misc : 3.47G/5ML/MSVOA W/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-9RE

Quant Time: Sep 24 07:28:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W092019S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 20 15:58:08 2019
Response via : Initial Calibration

