

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW092319\
 Data File : VW013239.D
 Acq On : 23 Sep 2019 18:39
 Operator : SY/VA
 Sample : K4957-03
 Misc : 5.01G/5ML/MSVOA W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SB-42-6.5-7.0

Quant Time: Sep 24 07:52:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W092019S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 20 15:58:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	160294	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	234797	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	195020	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	82992	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	66068	50.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.00%	
35) Dibromofluoromethane	7.88	113	65290	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
50) Toluene-d8	10.32	98	270979	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
62) 4-Bromofluorobenzene	12.62	95	80148	43.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.82%	
Target Compounds						
20) Methylene Chloride	4.90	84	12967	6.040	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW092319\
Data File : VW013239.D
Acq On : 23 Sep 2019 18:39
Operator : SY/VA
Sample : K4957-03
Misc : 5.01G/5ML/MSVOA W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SB-42-6.5-7.0

Quant Time: Sep 24 07:52:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W092019S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 20 15:58:08 2019
Response via : Initial Calibration

