


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW092319\
Data File : VW013244.D
Acq On : 23 Sep 2019 21:41
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W092019S.M
Title : SW846 8260
Last Update : Fri Sep 20 15:58:08 2019



AutoFind: Scans 1756, 1757, 1758; Background Corrected with Scan 1747

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.7	38517	PASS
75	95	30	60	47.6	103365	PASS
95	95	100	100	100.0	217280	PASS
96	95	5	9	6.8	14676	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	86.2	187328	PASS
175	174	5	9	8.2	15271	PASS
176	174	95	101	96.4	180565	PASS
177	176	5	9	7.3	13194	PASS