

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW092419\
 Data File : VW013285.D
 Acq On : 24 Sep 2019 19:05
 Operator : SY/VA
 Sample : K4657-04
 Misc : 4.83G/5ML/MSVOA W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BU-03-092419

Quant Time: Sep 25 08:14:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W092019S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 20 15:58:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	38118	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	55483	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	43827	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.55	152	12313	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	16741	53.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.62%	
35) Dibromofluoromethane	7.88	113	15818	51.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.68%	
50) Toluene-d8	10.32	98	60794	47.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.20%	
62) 4-Bromofluorobenzene	12.62	95	16241	37.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.46%	
Target Compounds						
16) Acetone	4.13	43	5205	74.192	ug/l #	80
20) Methylene Chloride	4.90	84	2822	5.266	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW092419\
Data File : VW013285.D
Acq On : 24 Sep 2019 19:05
Operator : SY/VA
Sample : K4657-04
Misc : 4.83G/5ML/MSVOA W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BU-03-092419

Quant Time: Sep 25 08:14:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W092019S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 20 15:58:08 2019
Response via : Initial Calibration

