


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW092419\
 Data File : VW013266.D
 Acq On : 24 Sep 2019 10:23
 Operator : SY/VA
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W092019S.M
 Title : SW846 8260
 Last Update : Fri Sep 20 15:58:08 2019



AutoFind: Scans 1756, 1757, 1758; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.1	40690	PASS
75	95	30	60	47.0	111688	PASS
95	95	100	100	100.0	237515	PASS
96	95	5	9	6.9	16482	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	88.2	209493	PASS
175	174	5	9	5.1	10764	PASS
176	174	95	101	96.8	202816	PASS
177	176	5	9	6.8	13892	PASS