

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW092518\
 Data File : VW005602.D
 Acq On : 25 Sep 2018 11:15
 Operator : VA/AP
 Sample : VW0925SBL01
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0925SBL01

Quant Time: Sep 25 14:12:03 2018
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W092118S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 21 09:05:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	220791	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	338362	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	309120	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	155753	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.31	65	126553	48.65	ug/l	0.00
Spiked Amount			Recovery	=	97.30%	
35) Dibromofluoromethane	7.89	113	107974	50.74	ug/l	0.00
Spiked Amount			Recovery	=	101.48%	
50) Toluene-d8	10.33	98	433732	50.06	ug/l	0.00
Spiked Amount			Recovery	=	100.12%	
62) 4-Bromofluorobenzene	12.63	95	150086	46.82	ug/l	0.00
Spiked Amount			Recovery	=	93.64%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW092518\
Data File : VW005602.D
Acq On : 25 Sep 2018 11:15
Operator : VA/AP
Sample : VW0925SBL01
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0925SBL01

Quant Time: Sep 25 14:12:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W092118S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 21 09:05:42 2018
Response via : Initial Calibration

