

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW092520\
 Data File : VW016638.D
 Acq On : 24 Sep 2020 10:18
 Operator : SY/VA
 Sample : VW0925SBL01
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK20

Manual Integrations
 APPROVED

MMDadoda
 9/28/2020 3:14:58 PM

Quant Time: Sep 25 08:08:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM091520S.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 25 07:55:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	607299	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	518001	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	246423	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.37	65	109946	18.70	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.80%
7) Chloroethane-d5	2.90	69	98696	20.56	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.24%
10) 1,1-Dichloroethene-d2	4.03	63	176733	13.45	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	53.80%
20) 2-Butanone-d5	7.10	46	57934	36.15	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	72.30%
24) Chloroform-d	7.65	84	293508	23.10	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.40%
26) 1,2-Dichloroethane-d4	8.31	65	148658	22.79	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.16%
29) Benzene-d6	8.28	84	526049	22.64	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	90.56%
33) 1,2-Dichloropropane-d6	9.28	67	150943	23.98	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	95.92%
37) Toluene-d8	10.33	98	479098	20.96	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	83.84%
38) trans-1,3-Dichloropropene-	10.58	79	67006	21.11	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.44%
39) 2-Hexanone-d5	10.93	63	45760	39.96	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.92%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	119783	22.36	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	89.44%
61) 1,2-Dichlorobenzene-d4	13.85	152	200922	25.52	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	102.08%

Target Compounds				Ovalue
16) Methylene chloride	4.94	84	9096m	0.971 ug/L

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW092520\
Data File : VW016638.D
Acq On : 24 Sep 2020 10:18
Operator : SY/VA
Sample : VW0925SBL01
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 4 Sample Multiplier: 1




Instrument :
MSVOA_W
ClientSampleId :
VBLK20

Manual Integrations
APPROVED

MMDadoda
9/28/2020 3:14:58 PM

Quant Time: Sep 25 08:08:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM091520S.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 25 07:55:30 2020
Response via : Initial Calibration





#16

Methylene chloride

Concen: 0.971 ug/L m

RT: 4.94 min Scan# 498

Delta R.T. 0.02 min

Lab File: VW016638.D

Acq: 24 Sep 2020 10:18

Tot Ion:	84	Resp:	9096
Ion Ratio	Lower	Upper	
84	100		
49	97.1	76.0	141.2
86	51.5	45.6	84.6

Instrument :
MSVOA_W
ClientSampleId :
VBLK20

Manual Integrations
APPROVED

MMDadoda
9/28/2020 3:14:58 PM

