

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW092718\
 Data File : VW005715.D
 Acq On : 27 Sep 2018 21:44
 Operator : VA/AP
 Sample : J4776-05RE
 Misc : 6.77G/5ML/MSVOA W/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EO-01-092718-ARE

Quant Time: Sep 28 02:35:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W092118S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 21 09:05:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	189014	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	308477	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	251355	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	92309	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	135760	60.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.94%	
35) Dibromofluoromethane	7.89	113	105918	54.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.20%	
50) Toluene-d8	10.33	98	374179	47.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.74%	
62) 4-Bromofluorobenzene	12.62	95	116521	39.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.74%	
Target Compounds						
16) Acetone	4.14	43	12903	28.499	ug/l	99
39) Methylcyclohexane	9.34	83	7239	1.976	ug/l	94
68) m/p-Xylenes	11.84	106	7309	2.043	ug/l	99
69) o-Xylene	12.17	106	6834	2.039	ug/l	98
80) 1,3,5-Trimethylbenzene	12.95	105	14353	2.579	ug/l	100
84) 1,2,4-Trimethylbenzene	13.26	105	37953	6.757	ug/l	95
95) Naphthalene	15.38	128	185786	57.171	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW092718\
Data File : VW005715.D
Acq On : 27 Sep 2018 21:44
Operator : VA/AP
Sample : J4776-05RE
Misc : 6.77G/5ML/MSVOA W/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EO-01-092718-ARE

Quant Time: Sep 28 02:35:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W092118S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 21 09:05:42 2018
Response via : Initial Calibration

