


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW092718\
Data File : VW005689.D
Acq On : 27 Sep 2018 10:00
Operator : VA/AP
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W092118S.M
Title : SW846 8260
Last Update : Fri Sep 21 09:05:42 2018



AutoFind: Scans 1758, 1759, 1760; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	24466	PASS
75	95	30	60	51.5	65973	PASS
95	95	100	100	100.0	128069	PASS
96	95	5	9	6.9	8839	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	85.4	109333	PASS
175	174	5	9	8.1	8863	PASS
176	174	95	101	96.6	105621	PASS
177	176	5	9	6.9	7296	PASS