




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092822\
Data File : VW024665.D
Acq On : 28 Sep 2022 18:01
Operator : SY/VA
Sample : N4852-10
Misc : 3.04g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW814

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 09/29/2022
Supervised By :Mahesh Dadoda 09/29/2022

Quant Time: Sep 29 04:46:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Sep 29 04:41:43 2022
Response via : Initial Calibration



TIC: VW024665.D\data.ms

(13) Acetone (T)

4.135min (+ 0.013) 6.45 ug/L

response 966

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	33.80	4.97
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092822\
 Data File : VW024665.D
 Acq On : 28 Sep 2022 18:01
 Operator : SY/VA
 Sample : N4852-10
 Misc : 3.04g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW814

Manual IntegrationsAPPROVED

Quant Time: Sep 29 04:46:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Sep 29 04:41:43 2022
 Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 09/29/2022
 Supervised By :Mahesh Dadoda 09/29/2022



TIC: VW024665.D\data.ms

(13) Acetone (T)

4.160min (+ 0.037) 13.92 ug/L m

response 2084

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	33.80	2.30
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092822\
 Data File : VW024665.D
 Acq On : 28 Sep 2022 18:01
 Operator : SY/VA
 Sample : N4852-10
 Misc : 3.04g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW814

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 09/29/2022
 Supervised By :Mahesh Dadoda 09/29/2022

Quant Time: Sep 29 04:46:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Sep 29 04:41:43 2022
 Response via : Initial Calibration

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene	8.842	114		54567	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117		48491	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152		19652	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3	2.355	65		9140	16.009	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150		Recovery =	64.040%		
7) Chloroethane-d5	2.892	69		12312	37.111	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150		Recovery =	148.440%		
11) 1,1-Dichloroethene-d2	4.019	63		16627	12.583	ug/L	0.00
Spiked Amount 25.000	Range 45	- 110		Recovery =	50.320%		
21) 2-Butanone-d5	7.086	46		10828	52.753	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135		Recovery =	105.500%		
24) Chloroform-d	7.653	84		31003	21.253	ug/L	0.00
Spiked Amount 25.000	Range 40	- 150		Recovery =	85.000%		
26) 1,2-Dichloroethane-d4	8.305	65		16602	21.939	ug/L	0.00
Spiked Amount 25.000	Range 70	- 130		Recovery =	87.760%		
32) Benzene-d6	8.281	84		52593	16.969	ug/L	0.00
Spiked Amount 25.000	Range 20	- 135		Recovery =	67.880%		
36) 1,2-Dichloropropane-d6	9.274	67		19167	20.240	ug/L	0.00
Spiked Amount 25.000	Range 70	- 120		Recovery =	80.960%		
41) Toluene-d8	10.323	98		40219	14.519	ug/L	0.00
Spiked Amount 25.000	Range 30	- 130		Recovery =	58.080%		
43) trans-1,3-Dichloroprop...	10.585	79		7199	17.071	ug/L	0.01
Spiked Amount 25.000	Range 30	- 135		Recovery =	68.280%		
47) 2-Hexanone-d5	10.927	63		6219	38.383	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135		Recovery =	76.760%		
56) 1,1,2,2-Tetrachloroeth...	12.689	84		17083	25.318	ug/L	0.00
Spiked Amount 25.000	Range 45	- 120		Recovery =	101.280%		
66) 1,2-Dichlorobenzene-d4	13.853	152		13163	19.035	ug/L	0.00
Spiked Amount 25.000	Range 75	- 120		Recovery =	76.120%		
Target Compounds							
13) Acetone	4.160	43		2084m	13.916	ug/L	
16) Methylene chloride	4.922	84		10949	10.217	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092822\
Data File : VW024665.D
Acq On : 28 Sep 2022 18:01
Operator : SY/VA
Sample : N4852-10
Misc : 3.04g/10mL/MSVOA_W/soil
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW814

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 09/29/2022
Supervised By :Mahesh Dadoda 09/29/2022

Quant Time: Sep 29 04:46:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Sep 29 04:41:43 2022
Response via : Initial Calibration

