

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092921\
 Data File : VW020223.D
 Acq On : 29 Sep 2021 11:23
 Operator : SY/VA
 Sample : M3973-04
 Misc : 7.74g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW8X0

Quant Time: Sep 30 01:02:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092121SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Sep 30 00:57:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	296111	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	260530	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	89345	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	103964	17.162	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	68.640%
7) Chloroethane-d5	2.898	69	91026	19.946	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.800%
11) 1,1-Dichloroethene-d2	4.025	63	116112	12.242	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	48.960%
21) 2-Butanone-d5	7.080	46	34503	36.910	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	73.820%
24) Chloroform-d	7.647	84	184039	19.763	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.040%
26) 1,2-Dichloroethane-d4	8.305	65	102610	20.473	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	81.880%
32) Benzene-d6	8.281	84	324500	19.523	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	78.080%
36) 1,2-Dichloropropane-d6	9.274	67	101796	21.162	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	84.640%
41) Toluene-d8	10.329	98	273646	18.094	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	72.360%
43) trans-1,3-Dichloroprop...	10.579	79	38429	18.691	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	74.760%
47) 2-Hexanone-d5	10.927	63	25975	39.020	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	78.040%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	74144	19.982	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	79.920%
66) 1,2-Dichlorobenzene-d4	13.853	152	73124	21.827	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	87.320%
Target Compounds					Qvalue	
13) Acetone	4.129	43	3458	4.319	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092921\
Data File : VW020223.D
Acq On : 29 Sep 2021 11:23
Operator : SY/VA
Sample : M3973-04
Misc : 7.74g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW8X0

Quant Time: Sep 30 01:02:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092121SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Sep 30 00:57:20 2021
Response via : Initial Calibration

