

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW093019\
 Data File : VW013347.D
 Acq On : 30 Sep 2019 22:38
 Operator : SY/VA
 Sample : VW0930SBL02
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0930SBL02

Quant Time: Oct 01 06:25:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W092019S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 20 15:58:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	320078	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	471719	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	397575	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.55	152	174993	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	135403	51.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.66%	
35) Dibromofluoromethane	7.88	113	135426	52.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.40%	
50) Toluene-d8	10.32	98	545007	50.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.38%	
62) 4-Bromofluorobenzene	12.62	95	171135	46.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.28%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW093019\
Data File : VW013347.D
Acq On : 30 Sep 2019 22:38
Operator : SY/VA
Sample : VW0930SBL02
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0930SBL02

Quant Time: Oct 01 06:25:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W092019S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 20 15:58:08 2019
Response via : Initial Calibration

