


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW093019\
 Data File : VW013345.D
 Acq On : 30 Sep 2019 20:53
 Operator : SY/VA
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W092019S.M
 Title : SW846 8260
 Last Update : Fri Sep 20 15:58:08 2019



AutoFind: Scans 1756, 1757, 1758; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.2	34474	PASS
75	95	30	60	46.7	93594	PASS
95	95	100	100	100.0	200618	PASS
96	95	5	9	7.0	14105	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	86.6	173760	PASS
175	174	5	9	8.0	13937	PASS
176	174	95	101	97.7	169677	PASS
177	176	5	9	7.2	12265	PASS