

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW093021\
 Data File : VW020262.D
 Acq On : 30 Sep 2021 11:55
 Operator : SY/VA
 Sample : M3877-15
 Misc : 5.08g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Oct 01 01:04:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092121SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Oct 01 01:03:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	386745	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	355444	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	163877	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	146330	18.495	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.960%
7) Chloroethane-d5	2.897	69	123397	20.703	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.800%
11) 1,1-Dichloroethene-d2	4.025	63	142630	11.514	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	46.040%
21) 2-Butanone-d5	7.086	46	53357	43.702	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.400%
24) Chloroform-d	7.653	84	298242	24.521	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.080%
26) 1,2-Dichloroethane-d4	8.305	65	164867	25.185	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	100.760%
32) Benzene-d6	8.275	84	580110	25.582	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	102.320%
36) 1,2-Dichloropropane-d6	9.274	67	165590	25.232	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.920%
41) Toluene-d8	10.323	98	537751	26.062	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	104.240%
43) trans-1,3-Dichloroprop...	10.579	79	70901	25.276	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	101.120%
47) 2-Hexanone-d5	10.926	63	40419	44.504	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	89.000%
56) 1,1,2,2-Tetrachloroeth...	12.694	84	115626	22.841	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	91.360%
66) 1,2-Dichlorobenzene-d4	13.853	152	132724	21.599	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	86.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW093021\
Data File : VW020262.D
Acq On : 30 Sep 2021 11:55
Operator : SY/VA
Sample : M3877-15
Misc : 5.08g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Oct 01 01:04:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092121SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Oct 01 01:03:15 2021
Response via : Initial Calibration

