

Quantitation Report (QT Reviewed)


Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW100118\
 Data File : VW005732.D
 Acq On : 01 Oct 2018 13:53
 Operator : SY/AP
 Sample : J5180-01
 Misc : 5.89G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampled :
 JLC10

Quant Time: Oct 02 06:29:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM092818S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 02 05:44:23 2018
 Response via : Initial Calibration

Manual Integrations
 APPROVED

sam
 10/5/2018 2:42:23 PM



Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW100118\
 Data File : VW005732.D
 Acq On : 01 Oct 2018 13:53
 Operator : SY/AP
 Sample : J5180-01
 Misc : 5.89G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampled :
 JLC10

Quant Time: Oct 02 05:44:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM092818S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 02 05:44:23 2018
 Response via : Initial Calibration

Manual Integrations
 APPROVED

sam
 10/5/2018 2:42:23 PM



(20) 2-Butanone-d5 (S)
 7.232min (+0.140) 0.06ug/L
 response 68

Ion	Exp%	Act%
46.00	100	100
77.00	25.60	29.41
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW100118\
 Data File : VW005732.D
 Acq On : 01 Oct 2018 13:53
 Operator : SY/AP
 Sample : J5180-01
 Misc : 5.89G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JLC10

Manual Integrations
 APPROVED

sam
 10/5/2018 2:42:23 PM

Quant Time: Oct 02 05:44:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM092818S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 02 05:44:23 2018
 Response via : Initial Calibration



TIC: VW005732.D

(20) 2-Butanone-d5 (S)

7.092min (+0.000) 35.64ug/L m > M010/05118

response 37808

Ion	Exp%	Act%
46.00	100	100
77.00	25.60	0.05#
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (QT Reviewed)

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW100118\
 Data File : VW005732.D
 Acq On : 01 Oct 2018 13:53
 Operator : SY/AP
 Sample : J5180-01
 Misc : 5.89G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 Client Sampled :
 JLC10

Manual Integrations
 APPROVED

sam
 10/5/2018 2:42:23 PM

Quant Time: Oct 02 06:29:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM092818S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 02 05:44:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev (Min)
1) 1,4-Difluorobenzene	8.85	114	259238	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	220738	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	78453	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	50619	15.32	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.28%
7) Chloroethane-d5	2.90	69	43842	17.00	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	68.00%
10) 1,1-Dichloroethene-d2	4.01	63	88266	12.23	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	48.92%
20) 2-Butanone-d5	7.09	46	37808m	35.64	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	71.28%
24) Chloroform-d	7.65	84	139620	20.96	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	83.84%
26) 1,2-Dichloroethane-d4	8.31	65	85944	21.31	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.24%
29) Benzene-d6	8.27	84	242744	22.21	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	88.84%
33) 1,2-Dichloropropane-d6	9.28	67	75420	23.88	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	95.52%
37) Toluene-d8	10.33	98	211317	19.59	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	78.36%
38) trans-1,3-Dichloropropene-	10.59	79	34321	20.69	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	82.76%
39) 2-Hexanone-d5	10.93	63	28092	37.98	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.96%
48) 1,1,2,2-Tetrachloroethane-	12.70	84	60496	19.72	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	78.88%
61) 1,2-Dichlorobenzene-d4	13.86	152	57155	20.59	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	82.36%

MD 10/05/18

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
16) Methylene chloride	4.91	84	12513	2.927	ug/L	95
53) Ethylbenzene	11.74	91	15782	0.989	ug/L	90
54) m,p-Xylene	11.84	106	41490	6.901	ug/L	99
55) o-xylene	12.17	106	19875	3.507	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed