

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW100219\
 Data File : VW013380.D
 Acq On : 02 Oct 2019 16:04
 Operator : SY/VA
 Sample : K5083-05
 Misc : 5.01G/5ML/MSVOA W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 20190704-RR-COMPOSITE-4

Quant Time: Oct 02 16:27:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100119S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 11:49:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	279619	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	417670	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	352057	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	158187	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	118627	56.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.64%	
35) Dibromofluoromethane	7.88	113	120096	54.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.60%	
50) Toluene-d8	10.32	98	476830	52.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.86%	
62) 4-Bromofluorobenzene	12.62	95	152014	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
Target Compounds						
16) Acetone	4.13	43	12307	19.231	ug/l	95
17) Carbon Disulfide	4.38	76	46792	6.516	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW100219\
Data File : VW013380.D
Acq On : 02 Oct 2019 16:04
Operator : SY/VA
Sample : K5083-05
Misc : 5.01G/5ML/MSVOA W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
20190704-RR-COMPOSITE-4

Quant Time: Oct 02 16:27:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W100119S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 11:49:57 2019
Response via : Initial Calibration

