

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW100219\
 Data File : VW013375.D
 Acq On : 02 Oct 2019 12:55
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 02 16:13:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100119S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 11:49:57 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.193	0.202	-4.7	113	0.00
3 P	Chloromethane	0.319	0.289	9.4	100	0.00
4 C	Vinyl Chloride	0.453	0.439	3.1#	105	0.00
5 T	Bromomethane	0.278	0.277	0.4	103	0.00
6 T	Chloroethane	0.262	0.245	6.5	103	0.00
7 T	Trichlorofluoromethane	0.196	0.197	-0.5	109	0.00
8 T	Diethyl Ether	0.227	0.213	6.2	101	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.436	0.423	3.0	103	0.00
10 T	Methyl Iodide	0.704	0.686	2.6	101	0.00
11 T	Tert butyl alcohol	0.037	0.028	24.3#	76	0.01
12 CM	1,1-Dichloroethene	0.460	0.451	2.0#	105	0.00
13 T	Acrolein	0.034	0.033	2.9	90	0.00
14 T	Allyl chloride	0.681	0.678	0.4	105	0.00
15 T	Acrylonitrile	0.103	0.091	11.7	98	0.00
16 T	Acetone	0.114	0.094	17.5	90	0.00
17 T	Carbon Disulfide	1.284	1.294	-0.8	104	0.00
18 T	Methyl Acetate	0.268	0.215	19.8	96	0.00
19 T	Methyl tert-butyl Ether	0.677	0.630	6.9	99	0.00
20 T	Methylene Chloride	0.532	0.463	13.0	101	0.00
21 T	trans-1,2-Dichloroethene	0.490	0.473	3.5	103	0.00
22 T	Diisopropyl ether	1.315	1.285	2.3	101	0.00
23 T	Vinyl Acetate	0.789	0.760	3.7	101	0.00
24 P	1,1-Dichloroethane	0.801	0.788	1.6	103	0.00
25 T	2-Butanone	0.146	0.124	15.1	92	0.00
26 T	2,2-Dichloropropane	0.514	0.474	7.8	107	0.00
27 T	cis-1,2-Dichloroethene	0.513	0.500	2.5	102	0.00
28 T	Bromochloromethane	0.290	0.316	-9.0	96	0.00
29 T	Tetrahydrofuran	0.085	0.074	12.9	96	0.00
30 C	Chloroform	0.785	0.748	4.7#	102	0.00
31 T	Cyclohexane	0.854	0.793	7.1	104	0.00
32 T	1,1,1-Trichloroethane	0.615	0.605	1.6	103	0.00
33 S	1,2-Dichloroethane-d4	0.373	0.381	-2.1	108	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.262	0.278	-6.1	109	0.00
36 T	1,1-Dichloropropene	0.464	0.456	1.7	104	0.00
37 T	Ethyl Acetate	0.218	0.182	16.5	96	0.00
38 T	Carbon Tetrachloride	0.402	0.409	-1.7	106	0.00
39 T	Methylcyclohexane	0.612	0.606	1.0	107	0.00
40 TM	Benzene	1.315	1.285	2.3	103	0.00
41 T	Methacrylonitrile	0.124	0.126	-1.6	109	0.00
42 TM	1,2-Dichloroethane	0.338	0.323	4.4	101	0.00
43 T	Isopropyl Acetate	0.386	0.350	9.3	98	0.00
44 TM	Trichloroethene	0.370	0.367	0.8	104	0.00
45 C	1,2-Dichloropropane	0.316	0.306	3.2#	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW100219\
 Data File : VW013375.D
 Acq On : 02 Oct 2019 12:55
 Operator : SY/VA
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 02 16:13:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100119S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 11:49:57 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.157	0.147	6.4	99	0.00
47 T	Bromodichloromethane	0.387	0.384	0.8	104	0.00
48 T	Methyl methacrylate	0.185	0.175	5.4	103	0.00
49 T	1,4-Dioxane	0.003	0.002	33.3#	98	0.00
50 S	Toluene-d8	1.089	1.164	-6.9	110	0.00
51 T	4-Methyl-2-Pentanone	0.203	0.177	12.8	98	0.00
52 CM	Toluene	0.848	0.830	2.1#	103	0.00
53 T	t-1,3-Dichloropropene	0.403	0.394	2.2	102	0.00
54 T	cis-1,3-Dichloropropene	0.488	0.483	1.0	103	0.00
55 T	1,1,2-Trichloroethane	0.235	0.221	6.0	102	0.00
56 T	Ethyl methacrylate	0.316	0.295	6.6	98	0.00
57 T	1,3-Dichloropropane	0.408	0.386	5.4	102	0.00
58 T	2-Chloroethyl Vinyl ether	0.130	0.128	1.5	101	0.00
59 T	2-Hexanone	0.140	0.125	10.7	93	0.00
60 T	Dibromochloromethane	0.268	0.266	0.7	102	0.00
61 T	1,2-Dibromoethane	0.227	0.213	6.2	99	0.00
62 S	4-Bromofluorobenzene	0.364	0.396	-8.8	114	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.380	0.364	4.2	101	0.00
65 PM	Chlorobenzene	1.024	1.001	2.2	102	0.00
66 T	1,1,1,2-Tetrachloroethane	0.349	0.352	-0.9	103	0.00
67 C	Ethyl Benzene	1.836	1.843	-0.4#	104	0.00
68 T	m/p-Xylenes	0.709	0.706	0.4	104	0.00
69 T	o-Xylene	0.655	0.662	-1.1	106	0.00
70 T	Styrene	1.120	1.132	-1.1	101	0.00
71 P	Bromoform	0.189	0.185	2.1	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	3.657	3.572	2.3	104	0.00
74 T	N-amyl acetate	0.778	0.730	6.2	102	0.00
75 P	1,1,2,2-Tetrachloroethane	0.619	0.560	9.5	103	0.00
76 T	1,2,3-Trichloropropane	0.444	0.439	1.1	101	0.00
77 T	Bromobenzene	0.878	0.841	4.2	105	0.00
78 T	n-propylbenzene	4.207	4.208	-0.0	108	0.00
79 T	2-Chlorotoluene	2.337	2.326	0.5	105	0.00
80 T	1,3,5-Trimethylbenzene	3.038	3.013	0.8	104	0.00
81 T	trans-1,4-Dichloro-2-butene	0.195	0.192	1.5	109	0.00
82 T	4-Chlorotoluene	2.448	2.429	0.8	108	0.00
83 T	tert-Butylbenzene	2.654	2.670	-0.6	106	0.00
84 T	1,2,4-Trimethylbenzene	3.020	2.996	0.8	107	0.00
85 T	sec-Butylbenzene	3.659	3.682	-0.6	108	0.00
86 T	p-Isopropyltoluene	3.391	3.431	-1.2	108	0.00
87 T	1,3-Dichlorobenzene	1.652	1.638	0.8	105	0.00
88 T	1,4-Dichlorobenzene	1.654	1.609	2.7	104	0.00
89 T	n-Butylbenzene	3.057	3.133	-2.5	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW100219\
Data File : VW013375.D
Acq On : 02 Oct 2019 12:55
Operator : SY/VA
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Oct 02 16:13:54 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100119S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 11:49:57 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.560	0.578	-3.2	107	0.00
91 T	1,2-Dichlorobenzene	1.461	1.421	2.7	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.093	0.082	11.8	98	0.00
93 T	1,2,4-Trichlorobenzene	1.048	1.046	0.2	102	0.00
94 T	Hexachlorobutadiene	0.681	0.688	-1.0	109	0.00
95 T	Naphthalene	1.768	1.725	2.4	102	0.00
96 T	1,2,3-Trichlorobenzene	0.915	0.907	0.9	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6