

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW100219\
 Data File : VW013396.D
 Acq On : 02 Oct 2019 22:58
 Operator : SY/VA
 Sample : K4946-04RE
 Misc : 3.77G/5ML/MSVOA W/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 GP-2 (1-3)RE

Quant Time: Oct 03 07:58:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100119S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 11:49:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	312154	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	451523	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	318209	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	85330	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	136486	58.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	117.12%	
35) Dibromofluoromethane	7.88	113	134716	56.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.72%	
50) Toluene-d8	10.32	98	505527	51.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.84%	
62) 4-Bromofluorobenzene	12.62	95	110183	33.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	67.12%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW100219\
Data File : VW013396.D
Acq On : 02 Oct 2019 22:58
Operator : SY/VA
Sample : K4946-04RE
Misc : 3.77G/5ML/MSVOA W/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
GP-2 (1-3)RE

Quant Time: Oct 03 07:58:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W100119S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 11:49:57 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.95 min Scan# 991

Delta R.T. -0.00 min

Lab File: VW013396.D

Acq: 02 Oct 2019 22:58

Instrument :

MSVOA_W

ClientSampleId :


GP-2 (1-3)RE

Tot Ion:168 Resp: 312154

Ion Ratio Lower Upper

168 100

99 48.2 40.2 60.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VW

7.95



150000

100000

50000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 58.558 ug/l

RT: 8.30 min Scan# 1049

Delta R.T. -0.01 min

Lab File: VW013396.D


Acq: 02 Oct 2019 22:58

Tgt Ion: 65 Resp: 136486

Ion Ratio Lower Upper

65 100

67 53.0 0.0 106.2



Abundance Ion 64.90 (64.60 to 65.60): VW

Ion 66.90 (66.60 to 67.60): VW

8.30


60000


40000

20000

0

Time-->





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.84 min Scan# 1138

Delta R.T. -0.00 min

Lab File: VW013396.D

Acq: 02 Oct 2019 22:58

Instrument :

MSVOA_W

ClientSampleId :

GP-2 (1-3)RE


Tot Ion:114 Resp: 451523

Ion Ratio Lower Upper

114 100

63 20.0 0.0 41.4




88 13.9 0.0 32.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VW

Ion 87.90 (87.60 to 88.60): VW



#35

Dibromofluoromethane

Concen: 56.860 ug/l

RT: 7.88 min Scan# 980

Delta R.T. -0.01 min

Lab File: VW013396.D

Acq: 02 Oct 2019 22:58


Tgt Ion:113 Resp: 134716

Ion Ratio Lower Upper

113 100

111 103.5 81.9 122.9



192 24.4 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 51.417 ug/l
RT: 10.32 min Scan# 1381
Delta R.T. -0.00 min
Lab File: VW013396.D
Acq: 02 Oct 2019 22:58




Instrument :
MSVOA_W
ClientSampleId :
GP-2 (1-3)RE


Tot Ion: 98 Resp: 505527
Ion Ratio Lower Upper
98 100
100 67.0 52.9 79.3



#62
4-Bromofluorobenzene
Concen: 33.559 ug/l
RT: 12.62 min Scan# 1757
Delta R.T. -0.00 min
Lab File: VW013396.D
Acq: 02 Oct 2019 22:58

Tgt Ion: 95 Resp: 110183
Ion Ratio Lower Upper
95 100
174 90.0 0.0 178.4
176 87.9 0.0 172.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.63 min Scan# 1595
Delta R.T. -0.01 min
Lab File: VW013396.D
Acq: 02 Oct 2019 22:58

Instrument :
MSVOA_W
ClientSampleId :
GP-2 (1-3)RE

Tot Ion:117 Resp: 318209
Ion Ratio Lower Upper
117 100
82 53.6 45.9 68.9
119 31.7 26.2 39.2




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.56 min Scan# 1912
Delta R.T. -0.00 min
Lab File: VW013396.D
Acq: 02 Oct 2019 22:58

Tgt Ion:152 Resp: 85330
Ion Ratio Lower Upper
152 100
115 55.2 27.3 81.9
150 157.1 0.0 349.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

