


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW100219\
 Data File : VW013374.D
 Acq On : 02 Oct 2019 12:19
 Operator : SY/VA
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100119S.M
 Title : SW846 8260
 Last Update : Wed Oct 02 11:49:57 2019



AutoFind: Scans 1756, 1757, 1758; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.6	35968	PASS
75	95	30	60	47.0	101650	PASS
95	95	100	100	100.0	216426	PASS
96	95	5	9	6.5	14090	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	84.2	182250	PASS
175	174	5	9	5.2	9462	PASS
176	174	95	101	99.2	180842	PASS
177	176	5	9	6.9	12502	PASS