

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW100220\
 Data File : VW016772.D
 Acq On : 02 Oct 2020 18:54
 Operator : SY/VA
 Sample : L4235-08
 Misc : 4.31G/10ML/MSVOA W/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AK4

Quant Time: Oct 03 06:20:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM091520S.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 03 04:48:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	585675	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	415481	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	100116	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.37	65	104015	18.35	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.40%
7) Chloroethane-d5	2.91	69	104614	22.60	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.40%
10) 1,1-Dichloroethene-d2	4.03	63	176365	13.92	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	55.68%
20) 2-Butanone-d5	7.09	46	48850	31.61	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	63.22%
24) Chloroform-d	7.66	84	298576	24.37	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.48%
26) 1,2-Dichloroethane-d4	8.31	65	143894	22.88	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.52%
29) Benzene-d6	8.28	84	566686	30.41	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	121.64%
33) 1,2-Dichloropropane-d6	9.28	67	160595	31.80	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	127.20%#
37) Toluene-d8	10.33	98	467921	25.53	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	102.12%
38) trans-1,3-Dichloropropene-	10.58	79	52291	20.54	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	82.16%
39) 2-Hexanone-d5	10.93	63	40317	43.90	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.80%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	93478	21.76	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.04%
61) 1,2-Dichlorobenzene-d4	13.86	152	77209	24.14	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.56%

Target Compounds					Ovalue
13) Acetone	4.14	43	2221	1.834 ug/L	56
16) Methylene chloride	4.93	84	11910	1.318 ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW100220\
Data File : VW016772.D
Acq On : 02 Oct 2020 18:54
Operator : SY/VA
Sample : L4235-08
Misc : 4.31G/10ML/MSVOA W/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AK4

Quant Time: Oct 03 06:20:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM091520S.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 03 04:48:41 2020
Response via : Initial Calibration

