

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW100318\
 Data File : VW005820.D
 Acq On : 03 Oct 2018 17:39
 Operator : SY/AP
 Sample : J5181-18
 Misc : 5.38G/10ML/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 JLC47

Quant Time: Oct 04 06:42:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM092818S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 04 05:16:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.85	114	212659	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	207125	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.57	152	88472	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	36230	13.36	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	53.44%
7) Chloroethane-d5	2.89	69	32388	15.31	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.24%
10) 1,1-Dichloroethene-d2	4.01	63	67474	11.40	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	45.60%
20) 2-Butanone-d5	7.08	46	46145	53.02	ug/L	-0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	106.04%
24) Chloroform-d	7.65	84	124946	22.87	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.48%
26) 1,2-Dichloroethane-d4	8.31	65	78997	23.87	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	95.48%
29) Benzene-d6	8.28	84	202393	19.74	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	78.96%
33) 1,2-Dichloropropane-d6	9.28	67	67124	22.65	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	90.60%
37) Toluene-d8	10.33	98	177649	17.55	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	70.20%
38) trans-1,3-Dichloropropene-	10.58	79	30445	19.56	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.24%
39) 2-Hexanone-d5	10.93	63	33462	48.22	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	96.44%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	71724	24.91	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	99.64%
61) 1,2-Dichlorobenzene-d4	13.86	152	72416	23.13	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	92.52%


Target Compounds					Qvalue
13) Acetone	4.13	43	6256	6.903 ug/L	97
16) Methylene chloride	4.92	84	36396	10.380 ug/L	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW100318\
Data File : VW005820.D
Acq On : 03 Oct 2018 17:39
Operator : SY/AP
Sample : J5181-18
Misc : 5.38G/10ML/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
JLC47

Quant Time: Oct 04 06:42:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM092818S.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 04 05:16:38 2018
Response via : Initial Calibration









#13
Acetone
Concen: 6.903 ug/L
RT: 4.13 min Scan# 365
Delta R.T. -0.01 min
Lab File: VW005820.D
Acq: 03 Oct 2018 17:39

Instrument :
MSVOA_W
ClientSampled :
JLC47

Tgt Ion: 43 Resp: 6256
Ion Ratio Lower Upper
43 100
58 17.0 0.0 31.8



#16
Methylene chloride
Concen: 10.380 ug/L
RT: 4.92 min Scan# 494
Delta R.T. -0.00 min
Lab File: VW005820.D
Acq: 03 Oct 2018 17:39

Tgt Ion: 84 Resp: 36396
Ion Ratio Lower Upper
84 100
49 120.1 87.2 162.0
86 65.5 43.2 80.2

