


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW100319\
Data File : VW013411.D
Acq On : 03 Oct 2019 09:51
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100119S.M
Title : SW846 8260
Last Update : Wed Oct 02 11:49:57 2019



AutoFind: Scans 1756, 1757, 1758; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.9	34722	PASS
75	95	30	60	47.3	97082	PASS
95	95	100	100	100.0	205151	PASS
96	95	5	9	6.8	13930	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	87.4	179242	PASS
175	174	5	9	5.2	9380	PASS
176	174	95	101	96.7	173333	PASS
177	176	5	9	6.9	12045	PASS