

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW100418\
 Data File : VW005841.D
 Acq On : 04 Oct 2018 07:29
 Operator : SY/AP
 Sample : VW1004SBL01
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW1004SBL01

Quant Time: Oct 04 09:18:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100418S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 04 07:08:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	168996	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	270947	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	269512	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	135535	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	113145	55.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.38%	
35) Dibromofluoromethane	7.89	113	92052	51.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.66%	
50) Toluene-d8	10.33	98	374280	53.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.16%	
62) 4-Bromofluorobenzene	12.63	95	129972	49.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.74%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW100418\
Data File : VW005841.D
Acq On : 04 Oct 2018 07:29
Operator : SY/AP
Sample : VW1004SBL01
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW1004SBL01

Quant Time: Oct 04 09:18:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W100418S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 04 07:08:57 2018
Response via : Initial Calibration

