

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW100419\
 Data File : VW013439.D
 Acq On : 04 Oct 2019 15:19
 Operator : SY/VA
 Sample : K5213-03
 Misc : 5.10G/5ML/MSVOA W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 S-2(0-5)

Quant Time: Oct 05 05:51:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W100119S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 11:49:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	306218	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	450368	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	362705	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	157614	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	124482	54.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.88%	
35) Dibromofluoromethane	7.88	113	128926	54.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.12%	
50) Toluene-d8	10.32	98	522639	53.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.58%	
62) 4-Bromofluorobenzene	12.62	95	169363	51.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.44%	
Target Compounds						
16) Acetone	4.14	43	9369	13.368	ug/l	93
27) cis-1,2-Dichloroethene	7.16	96	16340	5.200	ug/l	96
39) Methylcyclohexane	9.33	83	33115	6.011	ug/l	96
44) Trichloroethene	9.10	130	5220	1.565	ug/l	85
64) Tetrachloroethene	10.86	164	51869	18.797	ug/l	96
85) sec-Butylbenzene	13.38	105	31331	2.717	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW100419\
Data File : VW013439.D
Acq On : 04 Oct 2019 15:19
Operator : SY/VA
Sample : K5213-03
Misc : 5.10G/5ML/MSVOA W/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
S-2(0-5)

Quant Time: Oct 05 05:51:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W100119S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 11:49:57 2019
Response via : Initial Calibration

